



# Buddy lite<sup>®</sup>

BLOOD/FLUID WARMER

---

## OPERATOR'S MANUAL



# BUDDY LITE<sup>®</sup>

## BLOOD/FLUID WARMER

### OPERATOR'S MANUAL



All service calls & questions should be directed to:

USA: 855.397.4547  
Worldwide: +1.978.663.0212



Belmont Medical Technologies  
780 Boston Road  
Billerica, MA 01821, USA



EMERGO EUROPE  
WESTERVOORTSEDIJK 60  
6827 AT ARNHEM  
THE NETHERLANDS  
+31 (0) 70 345 8570

# Table of Contents

## CHAPTER 1: SYSTEM OVERVIEW

---

Introduction .....	1
Indications for Use and Contraindications .....	1
System Overview .....	2
System and Temperature Monitoring .....	3
Information Signals and Messages .....	3
Battery Pack .....	4

## CHAPTER 2: OPERATION

---

Introduction .....	5
Warnings .....	6
Step-by-Step Summary of Operating Procedures .....	7
Installing the Disposable Set .....	7
Connect to Fluid Bag and Administration Set .....	8
Prime the Disposable Set .....	8
Initiate Infusion .....	9
End of Procedure .....	10
Flow Characteristics .....	11
Operational Information Signals and Troubleshooting .....	12

## CHAPTER 3: SERVICE AND MAINTENANCE

---

Introduction .....	13
Cleaning, Disinfection, and Maintenance .....	14
Charging the Battery .....	15
Re-Calibrate the Fuel Gauge and Use of Alternate Charger .....	16,17
System Operational Checkout .....	18
Electromagnetic Compatibility .....	19

## CHAPTER 4: TECHNICAL SPECIFICATIONS

---

Technical Specifications .....	20
--------------------------------	----

CHAPTER 5: ORDERING INFORMATION .....	24
---------------------------------------	----

# System Overview

## INTRODUCTION:

---

The buddy lite<sup>®</sup> in-line Blood/Fluid warmer is a portable fluid warmer and sensing device that warms intra-venous fluids (including blood products) to physiological temperature at flow rates up to 80 ml/min (4.8 L/hr.) for crystalloids at 20°C and up to 50 ml/min (3.0 L/hr.) for packed red cells at 10°C. At higher flows, the fluids administered may not heat to normothermia.

The heater, including all control circuitry, is contained in a miniature package, and is powered by a small rechargeable battery. The whole device can be placed close to the patient to minimize line cooling in the infusion line. If the system overheats the fluid or is not heating or sensing temperature properly, the built-in red and blue LEDs will turn on alternately and the information signal sounds to alert the user of the device condition.

## INDICATIONS FOR USE AND CONTRAINDICATIONS

---

The buddy lite<sup>®</sup> Blood/Fluid Warmer is to be used to administer blood, blood products and intravenous solutions. It is intended to be used by healthcare professionals in a clinical environment to prevent hypothermia in children and adults. The system is not meant for the infusion of granulocyte suspensions.

## WARNING

The control system can only be used with the Belmont supplied disposable.

Do not alter the disposable, Heater Unit, or battery pack.

Use only with crystalloid or anticoagulated blood products.

Do not use in the presence of flammable anesthetics.

For use by trained professionals only. The buddy lite should not be left unattended while in operation.

## CAUTION

The system should not be used with pressure infusers. Blood and fluid bags contain air. The set can vent large amounts of air from crystalloid, but with blood only small amounts of air can be vented. The hand squeeze pumps, also known as a ball pump, supplied with some blood sets, may be used to momentarily increase flow, but use caution when doing so.

The system is an efficient blood and fluid warmer designed for use when the flow rate is 80 ml/min or less. It is neither a resuscitation device nor a substitute for a rapid infusion system.

## SYSTEM OVERVIEW

The buddy lite® Blood/Fluid Warmer complete transportable system consists of the battery pack, the **Heater Unit**, and the **Disposable Set**.

The system can be placed near the patient at the infusion site. See Figure 1.

The battery contains a built-in battery control circuitry and a built-in 5-segment LCD panel in the status window to provide the battery capacity. In cases when the alternate battery is used, capacity is indicated by 4 LEDs activated by a push button. Each LED represents 25% of the full charge capacity. When the system detects unsafe conditions, the built-in red and blue LEDs will turn on alternately and the information signal sounds to alert the user of the device condition.


The Heater Unit contains the heater plates and temperature monitoring and control system. If it is heating, the green LED will blink with each heating cycle.

The transportable system is for ground ambulances and helicopters with reciprocating and fixed wing engines.



**Figure 1: The buddy lite® Complete System**

**The disposable set** has a sterile fluid path and is **intended for single-use only**. The disposable set has standard luer connectors and can be connected to any standard catheter. It contains two thin films, which contact the heater plates for rapid heating of the fluid, and two internal microporous membranes to vent out-gassed air that is generated during heating, see Figure 2.



**Figure 2: The buddy® Disposable Set**

The set is color-coded and keyed to ensure proper placement in the Heater Unit, blue at the input from a fluid administration set. Both heater plates are heated during operation. There is a built-in valve in the disposable set to prevent air entrainment into the patient and a pressure-regulating valve, PRV, at the input to protect the set and the patient from unintended exposure to high pressure applied to the IV line. This valve will allow increase flow by application of pressure, up to 300 mmHg, but will prevent pressure higher than this from reaching the set or IV line distal to it. There is also a check valve at the output to prevent back flow.

## SYSTEM AND TEMPERATURE MONITORING

Blood or replacement fluid is heated as it passes through the disposable that is in contact with the heater plates in the Heater Unit. Thermistors on the heater boards measure the temperature of the heater plates, and a microprocessor, mounted on each of the heater plates, controls the temperature of the infusate. When the Heater Unit functions normally, the green LED on the Heater Unit will blink with each heating cycle. An over temperature condition causes the unit to stop heating, the green LED stops blinking, and an information signal will be activated, and the blue/red LED will flash alternately.


## INFORMATION SIGNALS AND MESSAGES:

Under information signal conditions, an audible information signal is activated and the blue/red LED flashes. Information signal conditions include: **Over Temperature, Probe Fault, Check for Air, and No Heat.**

## BATTERY PACK

---

The Battery Housing of the buddy lite<sup>®</sup> Blood/Fluid Warmer is shown below:



**Figure 3: Battery in the Battery Housing**

### POWER

The battery capacity is displayed via a 5-segment LCD panel on the battery pack. Each LCD segment represents 20% of the full charge capacity. If the battery voltage falls below 9.6 V, the system will shut down due to lack of power. In cases when alternate battery is used capacity is indicated by 4 LEDs activated by a push button. Each LED represents 25% of the full charge capacity.

The **ON/OFF switch** controls power from the battery pack to the Heater Unit.

### LED

There is a 2-color LED. The LED is **blue** when the battery power is ON and it flashes **red** when there is an information signal.

# Operation of the System

## INTRODUCTION

---

This section reviews the set up and operation of the buddy lite<sup>®</sup> Blood/Fluid Warmer to ensure safe and effective operation of the system.

**CAUTION:**

**Use of this device requires detailed familiarity with its operating procedures.**

**Read this manual thoroughly before operating this device.**



**WARNINGS:**

**Use only the Belmont supplied disposable. The disposable set is a single-use, sterile, non-pyrogenic fluid path.**

**Do not change or alter the disposable. Use extreme caution when handling the disposable; do not expose the disposable to sharp objects.**

**Do not use in the presence of flammable anesthetics.**

**Do not use in Oxygen rich environment.**

**Do not use this product in the presence of Nitrous Oxide.**

**Do not open the Heater Unit during or after priming.**

**Do not open the Heater Unit during operation, as this could damage the disposable.**

**Use only crystalloid to prime the unit. Do not use blood for the initial prime.**

**Do not attempt to prime unit or subject the disposable to flow outside of the Heater Unit, as the disposable will be damaged.**

**The patient line must be completely free of air before fluid is administered.**

**When administering blood products, be sure to use the administration set with a built-in filter.**

**Do not use for administration of cryo-precipitate, or granulocyte suspensions.**

**Use only anticoagulated blood products.**

**Do not mix lactated Ringer's or other solutions containing calcium with citrated blood products.**

**Practice standard hospital policy when handling blood products. Treat blood as if it were infected and clean up all spills immediately.**



**Under some Single Fault Conditions, the temperature of the top of the power supply enclosure could get hot and there is a possible risk of a burn if touched.**



**Do not submerge the buddy lite® system.**


**Store in a dust free environment and away from direct sunlight.**

**For use by trained professionals only. The buddy Lite should not be left unattended while in operation.**

## STEP-BY-STEP SUMMARY OF OPERATING PROCEDURES

Operating Steps	Detailed Information
<p><b>Set-up</b></p> 	<p>Make sure that you have all the necessary supplies.</p> <ul style="list-style-type: none"> <li>• Make sure that the battery is fully charged, 5-segment shown on the battery LCD or 4 LED if using alternate battery. If battery needs to be recharged, see the “Charging Procedure”, at the end of this section.</li> <li>• Inspect the Heater Unit to ensure that the heater plates are clean and dry. Wipe off any particles or debris on the plate with a soft cloth.</li> <li>• <b>There are no adjustments on the Heater Unit.</b></li> </ul>
<p><b>Heater Unit can be installed on top of the Battery Housing</b></p>	<p>The Heater Unit is placed close to the fluid insertion site during operation.</p>
<p><b>Install the Disposable Set</b></p> <p>Line up red orientation hub on the disposable with the notch in the Heater Unit, marked with red arrow (Red to Red). The Heater Unit will not close if they are not aligned.</p> 	<ul style="list-style-type: none"> <li>• Carefully remove the disposable from its pouch.</li> <li>• Install it into the Heater Unit, being careful to line up the red orientation hub on the disposable with the notch marked with a red arrow on the Heater Unit.</li> <li>• Take care not to damage the disposable.</li> </ul>

Operating Steps	Detailed Information
 <p style="text-align: center;">BLUE (INPUT)</p>	<p>The Heater Unit will not close if the disposable is not installed properly.</p>
 <p style="text-align: center;">Clamshell Latch</p>	<p>Close and latch the case.</p>
<p><b>Connect to the Fluid Bag and Administration Set</b></p>	<ul style="list-style-type: none"> <li>• Hang a fluid bag and administration set.</li> <li>• Connect the fluid administration set to the disposable luer fitting with the blue slide clamp.</li> </ul>
<p><b>Prime the Disposable Set</b></p> <p style="text-align: center;"><b>WARNING:</b></p> <p><b>The case must remain closed during and after priming.</b></p>	<ul style="list-style-type: none"> <li>• Open the roller clamps of the administration set and prime the unit making sure there is no air visible in the IV line or patient extension.</li> <li>• Tap the Heater Unit while priming to expel any air bubbles that may collect in the disposable.</li> </ul>

Operating Steps	Detailed Information
<p><b>Establish Power to the Heater Unit</b></p>	<p>Depress the ON/OFF switch, on the battery housing, to supply power to the Heater Unit.</p> <p>Blue LED turns ON and Red LED flashes once, if the Heater Unit is functioning normally. If not, return for service.</p>
<p><b>Initiate Infusion</b></p>  <p>The diagram illustrates the setup for initiating infusion. A central heater unit is connected to a patient via a line with a roller clamp. The patient connection is labeled 'MALE LUER' and 'PATIENT'. The fluid source connection is labeled 'FEMALE LUER' and 'CONNECT TO FLUID'. A 'BED SHEET CLAMP' is shown securing the line near the patient connection.</p>	<ul style="list-style-type: none"> <li>• Select an appropriate catheter size for the intended flow rate.</li> <li>• Using aseptic technique, make patient connection without entrapping air.</li> <li>• Adjust the roller clamp to the desired flow rate.</li> <li>• <b>Secure the system in a horizontal position, on a flat surface near the patient</b> using bed sheet clamp ensuring that there is no strain on the patient line.</li> <li>• If a longer administration set is desired, it should be no longer than 6 inches for optimal thermal transfer. Use of longer set will result in some line cooling from heater to patient.</li> <li>• The preferred position of the Heater Unit is at or below the level of the insertion site.</li> </ul>

Operating Steps	Detailed Information
<p><b>Maintain Infusion</b></p> <p><b>There are no adjustments on the Heater Unit.</b></p> <p>You may see some condensation around the outer edge of the Heater Unit after running for an extended period of time. This is normal and will not interfere with the system.</p> <p>If withdrawing fluid from distal end of disposable set via a syringe, do not exceed a rate of withdrawal of 100 ml/min.</p>	<ul style="list-style-type: none"> <li>• Observe the LED on the battery housing. The LED is BLUE when the battery power is ON.</li> <li>• The GREEN LED on the Heater Unit flashes with each heating cycle.</li> <li>• If there is an information signal, the BLUE and RED LEDs flash alternately. Respond to and correct system signals as needed.</li> </ul>
<p><b>End of Procedure</b></p>	<ul style="list-style-type: none"> <li>• Clamp roller clamp on the administration set off.</li> <li>• Clamp the input end of disposable set off using the blue slide clamp.</li> <li>• Turn Heater Unit off by pressing the ON/OFF switch on the battery housing.</li> <li>• Open the Heater Unit and remove disposable set.</li> <li>• Dispose of disposable set using hospital practices for bio-hazardous materials.</li> <li>• Remove the battery by pushing the tab holding the battery in the housing up and pull battery tab.</li> <li>• Recharge the battery.</li> </ul>

## FLOW CHARACTERISTICS

---

The buddy lite™ Blood/Fluid Warmer is designed for gravity flow. The maximum flow rates achievable will be determined by the size of the catheter employed.

The disposable has been designed to withstand up to 300 mmHg infusion pressure, provided the disposable set is within the Heater Unit and latched. The system is designed to maintain normothermic temperature at flow rates specified (50 ml/min for undiluted RBC's at 10°C, and 80 ml/min for 20° C crystalloid). However, at higher flows, the fluids administered may not heat to normothermia. **There are no adjustments for the Heater Unit.**

### **Pressure Drop:**

The disposable and heating unit does not restrict flow. If flow appears to be restricted:

- Check the infusion site to ensure catheter has not infiltrated the surrounding tissue
- Check to make sure line is not crimped or clamped off
- Out-gassed fluids may be causing air build up in the disposable. If air restriction is suspected:
  - Detach the unit from the patient connection
  - Holding the case with the patient connection port upward, tap the case gently while allowing a few seconds of free flow to remove air. Reattach the unit to the patient with a fluid-to-fluid connection.
  - If flow continues to be restricted, replace the disposable set.

### **Air Venting**

As fluid flows through the disposable set, it passes across microporous membranes, which can vent air from the fluid. The air is then released through the side vents of the set. **When infusing crystalloid, the set can vent large amounts of air. With blood or blood products, the set can vent only small amounts of air. CAUTION: When using blood or blood product, the Operator should be vigilant to avoid infusing air especially if using a hand squeeze pump during the infusion.** A pressure-regulating valve, incorporated into the disposable set, prevents abnormally high pressure greater than approximately 300 mmHg from reaching the set or the patient. A second valve built into the set prevents air entrainment by stopping flow if a partial vacuum is present.

During **normal operation** – water vapor, along with out-gassed air from warmed fluid, will be vented and will cause condensation to appear within the side vents after prolonged use. This should **not** be mistaken as a seal failure due to materials or workmanship.

## OPERATIONAL INFORMATION SIGNALS AND TROUBLESHOOTING

---

The chart below details information signals, the system response and the recommended user actions.

Probable Information Signal Condition	Condition and System Response	Probable Causes	Actions
<b>Check for Air</b>	System detects an empty set condition, flashes the red and blue LEDs alternately, and activates audible information signal. Heater continues to function normally.  System automatically resets and continues normal operation when air is cleared from the disposables.	Fluid source is empty or air in the fluid path.  Line is clamped shut.	If appropriate, replace fluid supply.  Open clamps.
<b>No Heat</b>	System detects no heat to the heater plates, activates audible information signal, and flashes the red and blue LEDs alternately.	Failure of the microprocessor or temperature sensor.	If failure reoccurs obtain replacement unit and contact service.
<b>Over Temperature</b>	System detects fluid temperature (i) Moderate to High Flow: >42.5°C for 2 min, or ≥ 48°C for 2.5 sec, or (ii) Low Flow: >42°C for 2.5 sec or ≥ 44°C for 1 sec, or (iii) Output Plate Temperature reaches 50.5°C. Stop heating, and flash red and blue LEDs alternately.	1. Repeatedly stopping and restarting during high flow conditions.  2. Failure of the microprocessor or temperature sensor.	Clamp line, turn unit off and re-power.  If failure reoccurs obtain replacement unit and contact service.
<b>Probe Fault</b>	Internal malfunction at start-up.	Hardware temperature override circuit not working properly.	Turn unit off and re-power. If not successful, call service.

**At these information signals, system stops heating, and flashes the red and blue LEDs alternately. First check for air in the system then reset or retry by pressing the ON/OFF switch to OFF, and then turn power to ON to resume operation.**

## Service and Maintenance

### INTRODUCTION

---

The buddy lite® Blood/Fluid Warmer has no internal user-serviceable parts and does not require calibration except for the battery fuel gauge. Routine maintenance consists of cleaning and inspecting the Heater Unit, recharging the battery after each use, recalibrating the battery fuel gauge, and testing the system operational.

**CAUTION**

The heater plates in the Heater Unit are sealed. Inspect the seals to ensure their integrity after each use. If the seal is not intact, do not use. Do not try to open or to service the Heater Unit.

**WARNING**

**Do not submerge the buddy lite® system.**

---



**ROUTINE MAINTAINENCE**

Turn the ON/OFF switch to OFF.

1. **Clean and/or Disinfect Exterior:**

**A. Battery Housing**

The housing may be wiped down with water or a 10% bleach solution using a soft cloth.

**Note:** Avoid the use of acetone or other solvents that may damage the surface.

**B. Heater Unit**

Wiping down the outside surfaces with a soft cloth moistened with water or a 10% bleach solution.

**Note:** Avoid the use of acetone or other solvents that may damage the surface.

Open the Heater Unit and inspect the seal around the heater plates. If the seal is intact, using a soft cloth wipe the heater plates with a 10% bleach solution and dry. If the seal around the heater plates is **not** intact, do not use and return the unit for service or replacement.

**Note:** Do not use abrasive cleaners on the Heater Unit or heater plates.

<p style="text-align: center;"><b>WARNING</b> <b>Do not submerge the Heater Unit.</b> <b>Practice standard hospital policy when handling blood products.</b> <b>Treat blood as if it were infected and clean up all spills immediately.</b></p>
---

2. **Visual Inspection:**

Once the unit is cleaned, inspect all wiring and connections for fraying or worn wiring. If noted, do not use; return the unit for service. If the wear is confined to the Heater Unit, replace the heater.

## CHARGING THE BATTERY:


---

(If using alternate charger, 50 Watt SMBUS desktop mini battery charger by RRC Power Solutions, refer to section 3.2 instruction)

### 1. Charger with Recalibrator

#### Material required:

- Charger/Calibrator
- 24V 2.5A DC power supply with universal mains input (90 to 260 VAC)
- Mains cable
  - 6' DC Cable



**Figure 4: Battery Charger and Calibrator**

#### Procedure:

- Place the “Charger and Calibrator” on a flat, level surface away from sources of heat and moisture.
- Plug the DC connector from the power supply into the back of the charger.
- Connect the power supply to the mains AC supply.
- Place the battery into the battery charger bay ensuring that the 5-way connector is fully seated.
- The LEDs in the “Charger Status Window” will provide status information and the charger will automatically begin charging the battery.

#### Charger LED Indication:

- **Green flashing:** Charging
- **Green solid:** Fully charged
- **Blue flashing:** In calibration mode
- **Blue solid:** Calibrated
- **Red flashing:** Need recalibration
- **Red solid:** Error

#### Charger LED Indication (12V)

- **Green flashing:** Charging
- **Green Solid:** Fully charged
- **Red solid:** Error

**Recharge Time:** It will take approximately 2.5 hours from fully discharged to fully charged.

**Recalibrate the Battery Fuel Gauge:**

If the battery fuel gauge needs to be recalibrated, the red LED indicator, on the battery charger, will flash. The recalibration is used to ensure the reliability of the battery capacity.

**Procedure:**

- a. Place the battery into the battery charger bay ensuring that the 5-way connector is fully seated.
- b. Press the “Calibration button”.
- c. The blue LED will flash to indicate that the battery is undergoing the recalibration cycle.
- d. When the blue LED is solid, the battery is fully charged and fully calibrated.
- e. Remove battery from the charger bay.

**2. Charger, Alternate**

**Material required:**

- Charger (RRC Power Solutions)
- 24V 2.5A DC power supply with universal mains input (90 to 260 VAC)
- Mains cable



**Figure 5: Battery Charger, Alternate**

Single bay charger

**Procedure:**

- a. Place the “Charger” on a flat, level surface away from sources of heat and moisture.
- b. Plug the DC connector from the power supply into the back of the charger.
- c. Connect the power supply to the mains AC supply.
- d. Place the battery into the battery charger bay ensuring that the 5-way connector is fully seated.
- e. The LEDs in the “Charger Status Window” will provide status information and the charger will automatically begin charging the battery.

**Charger LED Indication:**

- **One time Red/Orange/Green:** Self-test: Charger is ready for use
- **Red/Green blinking:** Battery recognition and initialization
- **Orange light:** Battery is correct type and is being charged
- **Green light:** Battery is charged and can be removed for use
- **Red blinking:** Temperature limit.
  - **Too cold:** Battery will be charged when warmed sufficiently.
  - **Too hot:** Battery should be removed to cool down.
- **Red light:** The battery is damaged or it is a conventional battery which cannot be recharged.

**Recharge Time:** It will take approximately 2.75 hours from fully discharged to fully charged.

## **SYSTEM OPERATIONAL CHECKOUT**

This system operational checkout including the information signal test should be done periodically. Prior to performing the test, have the following equipment and supplies on hand:

- Saline or other crystalloid for testing
- Disposable Set
- Administration set
- Graduated cylinder and timing device for measuring flow rate

### **Setup:**

1. Make sure the battery is fully charged.
2. Install disposable set. The Heater Unit will not close if the disposable set is not installed properly. Close and latch the case.

### **Connect to the fluid bag and administration set:**

1. Hang a fluid bag and administration set. Connect the fluid administration set to the disposable luer fitting with blue slide clamp. Connect administration set with 14-gauge catheter to the output luer fitting.
2. Open the roller clamps and prime the unit. Tap the Heater Unit while priming to expel any air bubbles that may collect in the disposable.
3. Secure the Heater Unit on the battery housing, making sure that the output of the catheter is at least 5" higher than the Heater Unit to simulate a venous pressure of approximately 10 mmHg.

### **Operational Test:**

1. Depress the ON/OFF switch on the battery housing to turn the battery power to ON. The Blue LED should turn on and the Red LED flashes once. If not, return the Heater Unit for service.
2. Set the flow rate to approximately 80 ml/min.
3. Verify that the GREEN LED, inside the Heater Unit, flashes with each heating cycle.
4. Measure the output temperature at the end of the catheter. It should be  $38^{\circ} \pm 2^{\circ}\text{C}$ , with no information signal.
5. Turn the Heater Unit off by pressing the ON/OFF switch on the battery housing. Discard the disposable set.

## ELECTROMAGNETIC COMPATIBILITY

The buddy lite® is intended for use in the electromagnetic environment specified below. You should assure that it is used in such environment.

Test	Compliance	Electromagnetic environment - guidance
RF emissions, CISPR 11	Group 1, Class B	The buddy lite® is a battery operated Blood/Fluid Warmer. RF emission is negligible at any single frequency.
RS103 Radiated E-Field Susceptibility	RS103 of MIL-STD 461F for (Aircraft Internal)	The buddy lite™ was tested and passed in accordance with the radiated susceptibility requirements to RS103 of MIL-STD 461F for Aircraft Internal applications.
Electrostatic discharge (ESD) IEC 61000-4-2	± 6 kV contact ± 8 kV air	Floors should be wood, concrete or ceramic tile. If floors are covered with synthetic material, the relative humidity should be at least 30%.
Radiated RF IEC 61000-4-3	3 V/m over the frequency range of 80 MHz to 2.5 GHz	Radiated RF frequency is negligible at any single frequency.
Power Frequency Magnetic Field Immunity IEC 61000-4-8	3 A/m	If interference occurs, it may be necessary to position the buddy lite® further from sources of power frequency magnetic field.

## Technical Specifications

<b>DIMENSIONS</b>	
Size	Battery Housing: 3.33" x 7.26" x 1.36" (W x L x H) Battery: 3.1" x 5.9" x 0.89" Heater Unit: 1.5" x 5.2" x 0.87"
Weight	Total Weight: 1.6 lb Battery Housing: 0.28 lb Battery: 0.97 lb Heater Unit: 0.22 lb
<b>POWER</b>	
Battery Nominal Voltage	14.4 VDC
Electrical Compliance	EN 60601-1, UL60601-1, CAN/CSA C22.2-No. 601.1-M90
<b>ENVIRONMENTAL</b>	
Transportable Use (No fixturing needed on battery)	Carrying, stretcher, trolley, cart, road ambulance Helicopter: Reciprocating and Turbojet Engines
Operating Temperature	10°C to 40°C
Operating Relative Humidity	15% to 90% RH
Operating Pressure	62kPa to 106 kPa
Storage Temperature <sup>1</sup>	-20°C to 45°C
Warm-Up/ Cool-Down from Storage Extremes	2 minutes



Storage Relative Humidity	15% to 90% RH
Transient Use Temperature	10°C to 40°C
Transient Use Humidity	15% to 90% RH
Transient Use Pressure	62 to 106 kPa
Shock and Vibration	ISTA Pre-shipment Test Procedure, Test procedure 1, 1996
Electromagnetic Compatibility	EN 60601-1-2









<sup>1</sup> If stored at very low temperatures for extended periods of time, the battery charge may decrease.

<b>OPERATING PARAMETERS</b>	
Flow Rate	Input temperature 20°C: Max 80 ml/min Input temperature 10°C: Max 50 ml/min
Output Temperature:	38° ± 2°C
Prime Volume	4 ml without tubing, 7.5 ml with valves and tubing at the input and output
<b>OPERATING PANEL</b>	
Control Keys (On Battery Housing)	On/off Switch – Turn battery ON/OFF
LED (On Battery Housing)	BLUE: Solid – Battery power is ON  RED: Flashing once at start-up – Heater Unit is working  RED: Flashing alternate with BLUE flashing - Information Signal
LED (On Heater Unit)	GREEN: Flashing with each heating cycle.
<b>SAFETY AND MONTORING</b>	
Infusate Temperature	Heater power shut off if fluid temperature exceeds specified limits



Independent Safety Circuit	Shut off the heater if temperature exceeds a limiting value
<b>Information Signal</b>	CHECK FOR AIR NO HEAT OVER TEMPERATURE PROBE FAULT
<b>DISPOSABLE SET</b>	
Disposable Set	Sterile, Non-Pyrogenic Fluid Path, Single-Patient Use Only
Sterilization Method	Ethylene Oxide
Sterility Assurance Level	Greater than or equal to 10 <sup>-6</sup>

<b>CLASSIFICATION</b>	
Type of Protection Against Electrical Shock	Class II
Degree of Protection Against Electric Shock	CF defibrillator-proof
Degree of Protection Against Harmful Ingress of Water	IPX54, protected against water splashes
Degree of Safety in Presence of Flammable Anesthetics	Not suitable
Mode of Operation	Continuous
<b>Medical Equipment</b>  	Medical – General Medical Equipment

SYMBOLS AND DEFINITIONS	
Symbol	Description
	Compliance to Medical Device Directive 93/42/EEC
  or	Attention, consult accompanying documents/refer to manual
	Defibrillator-proof type CF equipment
<b>IPX54</b>	Protected against 1 mm or larger objects and water splashes
<b>SN</b>	Serial Number
	Do not immerse into water or any other liquid
	Manufactured by
	Authorized European Representative
	Waste Electrical and Electronic Equipment

### Waste Electrical and Electronic Equipment (WEEE)

The crossed-out wheel bin symbol on the product, literature, or packaging reminds you that all electrical and electronic products, batteries, and accumulators must be taken to separate collection at the end of their working life. This requirement applies to the European Union and other locations where separate collection systems are available. To prevent possible harm to the environment or human health from uncontrolled waste disposal, please do not dispose of these products as unsorted municipal waste, but instead, hand in at an official collection point for recycling.

## Ordering Information

### **Buddy lite® with single battery and charger**

buddy lite® Blood/Fluid Warmer System, N. America	P/N 905-00017
buddy lite® Blood/Fluid Warmer System, European	P/N 905-00016
buddy lite® Blood/Fluid Warmer System, Germany	P/N 905-00016-G
buddy lite® Blood/Fluid Warmer System, UK	P/N 905-00018
buddy lite® Blood/Fluid Warmer System, Switzerland	P/N 905-00019
buddy lite® Blood/Fluid Warmer System, Netherlands	P/N 905-00029
buddy lite® Blood/Fluid Warmer System, Sweden	P/N 905-00038
buddy lite® Blood/Fluid Warmer System, Australia	P/N 905-00045
buddy lite® Blood/Fluid Warmer System, China	P/N 905-00050
buddy lite® Blood/Fluid Warmer System, N. America, 12V Charger	P/N 905-00042
buddy lite® Blood/Fluid Warmer System, European, 12V Charger	P/N 905-00043
buddy lite® Blood/Fluid Warmer System, UK, 12V Charger	P/N 905-00044

### **Accessories**

Battery Pack	P/N 101-00029
Battery Charger, N. America	P/N 111-00026
Battery Charger, European	P/N 111-00027
Battery Charger, UK	P/N 111-00028
12V Battery Charger, N. America	P/N 111-00032
12V Battery Charger, European	P/N 111-00033
12V Battery Charger, UK	P/N 111-00034

### **Disposables**

buddy™ Disposable set	P/N 905-00010
-----------------------	---------------